

# ASEPT®

## 600ml or 1,000ml

### Bare Evacuated Drainage Bottle

#### INSTRUCTIONS FOR USE



LS-00165-01-AB REV 2016-04

WARNING: This product contains chemicals known to the State of California to cause cancer and birth defects or other reproductive harm.

Store at room temperature

This product contains dry natural rubber (latex)



Do not reuse



Do not resterilize



Do not use if package is damaged



Manufacturer



Keep away from sunlight



Consult instructions for use

**REF**

Catalog number

**LOT**

Batch code



Sterilized using ethylene oxide, sterile fluid pathway



Keep dry



Use by



Not made with DEHP

**Rx only**

Interventional Systems  
**B | BRAUN**

Distributed by:  
**B. Braun Interventional Systems Inc.**  
824 Twelfth Avenue  
Bethlehem, PA 18018  
www.bisusa.org



Manufacturer:  
pfm medical, inc.  
1815 Aston Ave.,  
Suite 106  
Carlsbad, CA 92008

Distributed by:  
**B. Braun Interventional Systems Inc.**  
824 Twelfth Avenue  
Bethlehem, PA 18018  
www.bisusa.org

Customer Service, ordering  
TEL: (877) VENA-CAV (836-2228)  
FAX: 1-(610)-849-1334  
Technical Support  
TEL: (800) 443-VENA (8362)  
Made in USA or Germany

Interventional Systems  
**B | BRAUN**

## Instructions for Use

### **ASEPT® 600ml or ASEPT® 1,000ml Bare Evacuated Drainage Bottle**

Disposable - This device is intended for one use only.  
Do not reuse or resterilize. Sterilized with ethylene oxide, sterile fluid path.

#### **PRODUCT DESCRIPTION:**

The ASEPT® 600ml or 1,000ml Bare Evacuated Drainage Bottle is a vacuum based suction device for drainage purposes. The bottle comes with a needle-free valve attached to the bottle. The following instructions are for the ASEPT® 600ml or 1,000ml Bare Evacuated Drainage Bottle.

#### **The ASEPT® 600ml or ASEPT® 1,000ml Bare Evacuated Drainage Bottle contains the following items:**

##### **One 600ml or 1,000ml vacuum bottle with the following features:**

- Bottle scaled in 50ml increments
- Green vacuum indicator
- Vacuum release clamp
- Needle-free valve with protective cap

#### **CAUTIONS:**

- The green indicator should be down and level with the "max" line; this indicates that the inside of the bottle contains a vacuum
- Before use, make sure the protective cap is on the needle-free valve
- When the green vacuum indicator reaches the position "min" a change of bottle is necessary
- Federal Law (USA) restricts this device to sale by or on the order of a physician

#### **INDICATION:**

The ASEPT® Drainage Bottle is a non-powered, single patient, portable suction device used to aspirate, remove, or sample body fluids.

#### **APPLICATION:**

Using the ASEPT® 600ml or 1,000ml Bare Evacuated Drainage Bottle.

Follow these step-by-step directions.

1. Pick up the ASEPT® 600ml or 1,000ml Bare Evacuated Drainage Bottle. Look at the green vacuum indicator on the top of the bottle. This indicator tells you if there is a vacuum in the bottle. If the indicator is up and level with the "min" line or the protective cap is missing then discard the bottle and use another.
2. Take the protective cap off the needle-free valve, swab the needle-free valve with a disinfectant and use a luer connector to connect to the needle-free valve. Twist all the way down ensuring the connection is secure.
3. To start fluid collection, slide the white slide clamp on the green tubing until it no longer pinches the green tube closed.
4. After the drainage procedure, disconnect the drain line from the needle-free valve.
5. The needle-free valve may be used for fluid sampling. To sample fluid, swab the needle-free valve with a disinfectant and use a needle-free syringe to access the valve, hold the bottle upside down and remove the fluid.
6. Dispose of the used ASEPT® 600ml or 1,000ml Bare Evacuated Drainage Bottle according to facility policies and procedures.

#### **STERILITY:**

This device has been sterilized, is fluid path sterile, is for single use only, and is not to be reused. B. Braun Interventional Systems will not be responsible for any products that are resterilized, nor accept for exchange or credit any product that has been altered but not used by the patient or purchaser.

**WARRANTY:** B. BRAUN INTERVENTIONAL SYSTEMS INC. WARRANTS THAT THIS PRODUCT WAS MANUFACTURED ACCORDING TO APPLICABLE STANDARDS AND SPECIFICATIONS. PATIENT CONDITION, CLINICAL TREATMENT AND PRODUCT MAINTENANCE MAY AFFECT THE PERFORMANCE OF THIS PRODUCT. USE OF THIS PRODUCT SHOULD BE IN ACCORDANCE WITH THE INSTRUCTIONS PROVIDED AND AS DIRECTED BY THE PRESCRIBING PHYSICIAN. B. BRAUN INTERVENTIONAL SYSTEMS INC. SHALL NOT BE LIABLE FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES OF ANY KIND.